

# ISO 17025 Compliant Calibration Certificate

14A0054A-IC-6788-230111-1

Oxford Technical Solutions



4647



## Table of Contents

1 Overview .....	1
1.1 Calibration Information .....	2
1.2 Certificated Item .....	2
2 Physical Conditions .....	2
2.1 Local Gravity Estimation .....	2
3 Equipment Used .....	2
3.1 Frequency Meter .....	2
3.2 Voltage Supply .....	3
3.3 Temperature Sensor .....	3
3.4 Calibration Table .....	3
3.5 Calibration Software .....	3
4 Environmental Conditions .....	3
5 Calibration Validation .....	4
5.1 Environmental Results .....	4
5.2 Accelerometer Results .....	4
5.3 Gyroscope Results .....	5
5.4 Summaries .....	7
6 Certificate Authorization .....	7

This calibration certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurements to the SI system of units and/or to units of measurement realised of the National Physical Laboratory or other recognised institutes.

This calibration has been performed in compliance with the requirements of ISO/IEC 17025:2017 (BS EN ISO/IEC 17025:2017). Oxford Technical Solutions is a UKAS accredited calibration laboratory No. 4647.

This certificate states the performance of the product after any change to the internal IMU sensor model.

Each reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor of two, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Other measurements made by the product are not traceable.

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

## 1. Overview

Issued by

Oxford Technical Solutions Ltd.

77 Heyford Park, Upper Heyford, Oxfordshire, OX25 5HD, UK  
Registered in England and Wales No. 3534778

UKAS accreditation number 4647  
End-user [REDACTED]  
Consignee [REDACTED]  
Confidentiality Confidential customer information

### 1.1. Calibration Information

Calibration by [REDACTED]  
Calibration ID [REDACTED]  
Calibration method 14A0054A  
Calibration date 2023-01-11  
Document revision 1

### 1.2. Certificated Item

This calibration only applies to the product listed here:

Model RT3003GG  
Serial number [REDACTED]  
Result Pass  
Condition Old

This product generates status information giving indications of the accuracy of each of the outputs. This information should be monitored to ensure that the outputs are within the accuracy required.

## 2. Physical Conditions

Certified fixed physical conditions used during the calibration:

### 2.1. Local Gravity Estimation

Name Estimated Value Of Gravity  
Calibration by British Geological Survey  
Calibration ID RG/84/14  
Calibration date 2005-11-28  
Location 78 Heyford Park, Upper Heyford, Oxfordshire, OX25 5HD, UK  
Estimation (m/s<sup>2</sup>) 9.81204 ± 0.00001

## 3. Equipment Used

Equipment used during the calibration:

### 3.1. Frequency Meter

Name Thurlby Thandar TF930

Serial	444273
Calibration by	RS Calibration
Calibration ID	1756790
Calibration date	2022-08-25
Calculated ratiometric uncertainty	0.0000040
Frequency uncertainty (Hz)	0.0004

### 3.2. Voltage Supply

Name	Fluke 289 Digital Multimeter
Serial	55550087
Calibration by	RS Calibration
Calibration ID	1778408
Calibration date	2022-12-28
Absolute uncertainty (V)	0.05

### 3.3. Temperature Sensor

Name	Digitron 2000T
Serial	5604251717
Calibration by	C.I.S Calibration Laboratories
Calibration ID	68466
Calibration date	2022-09-20
Absolute uncertainty (°C)	1.21
Difference uncertainty (°C)	0.25
Measurement bias (°C)	0.29

### 3.4. Calibration Table

Name	OxTS Calibration Table
Model	14A0052A
Serial	2

### 3.5. Calibration Software

Name	OxTS Calibrate
Version	220412.14g

## 4. Environmental Conditions

Recorded environmental conditions during the calibration:

IMU frequency (Hz)	100.000 ± 0.0004
--------------------	------------------

Start nominal voltage (V)	15.2 ± 0.05
Final nominal voltage (V)	15.2 ± 0.05
Start temperature (°C)	35.8 ± 1.21
Final temperature (°C)	36.4 ± 1.21

## 5. Calibration Validation

Four categories of result are used when presenting measurements:

- Pass     The result can be classified as a pass with a confidence in excess of 95%
- Pass \*   The error has a probability of between 50% and 95% of being inside the limit
- Fail \***   The error has a probability of between 5% and 50% of being inside the limit
- Fail**     The result can be classified as a fail with a confidence in excess of 95%

The measurement process used during the calibration of the product causes the uncertainties to vary: some measurements are more accurate than others along specific sensor axes; this is not a fault in the product, but an unavoidable limitation of the calibration process.

### 5.1. Environmental Results

The detailed analysis of temperature is shown in the following table, where

- Change = Final Temperature - Start Temperature
- Result = |Change| ± Uncertainty < Limit

**Table 1. Analysis of temperature conditions**

Change °C	Uncertainty °C	Limit °C	Result
0.60	0.25	1.00	Pass

### 5.2. Accelerometer Results

The detailed analysis for the three axis accelerations is shown in the following tables, where

- Error = Output acceleration - Input acceleration
- Limit = 0.01 m/s<sup>2</sup>
- Result = |Error| ± Uncertainty < Limit

**Table 2. Analysis of the x-axis acceleration**

Input m/s <sup>2</sup>	Output m/s <sup>2</sup>	Error m/s <sup>2</sup>	Uncertainty m/s <sup>2</sup>	Result
-9.4948	-9.4947	0.0000	0.0010	Pass
-8.1602	-8.1605	-0.0002	0.0010	Pass
-6.8600	-6.8599	0.0001	0.0010	Pass
-4.9594	-4.9595	-0.0001	0.0013	Pass
-4.4556	-4.4555	0.0001	0.0014	Pass
-3.2878	-3.2878	0.0000	0.0018	Pass
-1.1749	-1.1748	0.0001	0.0049	Pass
-0.3061	-0.3060	0.0002	0.0062	Pass
0.2085	0.2087	0.0002	0.0062	Pass
2.6979	2.6981	0.0002	0.0022	Pass

Input m/s <sup>2</sup>	Output m/s <sup>2</sup>	Error m/s <sup>2</sup>	Uncertainty m/s <sup>2</sup>	Result
4.4504	4.4507	0.0002	0.0014	Pass
5.0675	5.0678	0.0003	0.0013	Pass
7.1042	7.1045	0.0003	0.0010	Pass
8.1335	8.1339	0.0004	0.0010	Pass
9.4982	9.4983	0.0001	0.0010	Pass

**Table 3. Analysis of the y-axis acceleration**

Input m/s <sup>2</sup>	Output m/s <sup>2</sup>	Error m/s <sup>2</sup>	Uncertainty m/s <sup>2</sup>	Result
-9.5259	-9.5261	-0.0002	0.0010	Pass
-8.3464	-8.3467	-0.0003	0.0010	Pass
-6.7624	-6.7623	0.0000	0.0010	Pass
-4.9774	-4.9775	-0.0000	0.0013	Pass
-3.2580	-3.2580	0.0000	0.0018	Pass
-2.8081	-2.8081	0.0000	0.0021	Pass
-1.3826	-1.3825	0.0001	0.0042	Pass
-0.0212	-0.0210	0.0002	0.0062	Pass
2.2586	2.2590	0.0004	0.0026	Pass
2.8160	2.8164	0.0004	0.0021	Pass
3.2532	3.2535	0.0003	0.0019	Pass
4.9779	4.9783	0.0004	0.0013	Pass
6.8720	6.8726	0.0006	0.0010	Pass
8.3550	8.3554	0.0004	0.0010	Pass
9.5331	9.5336	0.0004	0.0010	Pass

**Table 4. Analysis of the z-axis acceleration**

Input m/s <sup>2</sup>	Output m/s <sup>2</sup>	Error m/s <sup>2</sup>	Uncertainty m/s <sup>2</sup>	Result
-9.8072	-9.8073	-0.0001	0.0010	Pass
-8.3528	-8.3533	-0.0005	0.0010	Pass
-5.1469	-5.1469	-0.0000	0.0012	Pass
-5.1428	-5.1428	-0.0000	0.0012	Pass
-4.6244	-4.6245	-0.0001	0.0014	Pass
-4.6162	-4.6164	-0.0002	0.0014	Pass
-0.2763	-0.2763	0.0001	0.0062	Pass
-0.2723	-0.2721	0.0002	0.0062	Pass
0.3011	0.3012	0.0001	0.0062	Pass
4.6407	4.6412	0.0005	0.0014	Pass
4.6440	4.6446	0.0006	0.0014	Pass
5.1518	5.1523	0.0005	0.0012	Pass
8.6533	8.6539	0.0006	0.0010	Pass
8.6544	8.6550	0.0006	0.0010	Pass
9.8077	9.8077	0.0001	0.0010	Pass

### 5.3. Gyroscope Results

The detailed analysis for the three axis angular rates is shown in the following tables, where

Error = Output angular rate - Input angular rate

Limit = 0.12 °/s

Result = |Error| ± Uncertainty < Limit

**Table 5. Analysis of the x-axis angular rate**

Input %/s	Output %/s	Error %/s	Uncertainty %/s	Result
-23.805	-23.810	-0.005	0.039	Pass
-19.158	-19.163	-0.005	0.042	Pass
-11.167	-11.180	-0.013	0.026	Pass
-7.878	-7.892	-0.013	0.019	Pass
-6.416	-6.426	-0.010	0.036	Pass
-0.011	-0.020	-0.010	0.037	Pass
-0.006	-0.016	-0.010	0.040	Pass
-0.001	-0.007	-0.006	0.041	Pass
0.006	0.001	-0.005	0.039	Pass
0.011	0.002	-0.009	0.037	Pass
4.377	4.364	-0.013	0.021	Pass
6.536	6.533	-0.003	0.037	Pass
9.698	9.683	-0.016	0.032	Pass
14.847	14.830	-0.017	0.031	Pass
21.786	21.769	-0.018	0.034	Pass

**Table 6. Analysis of the y-axis angular rate**

Input %/s	Output %/s	Error %/s	Uncertainty %/s	Result
-22.380	-22.380	-0.000	0.037	Pass
-17.378	-17.379	-0.000	0.038	Pass
-10.715	-10.702	0.013	0.018	Pass
-5.805	-5.801	0.004	0.020	Pass
-2.590	-2.586	0.004	0.019	Pass
-0.009	0.000	0.009	0.038	Pass
-0.001	-0.000	0.001	0.028	Pass
-0.000	0.003	0.003	0.039	Pass
0.001	0.004	0.003	0.035	Pass
0.009	0.014	0.004	0.030	Pass
5.207	5.208	0.001	0.035	Pass
7.781	7.775	-0.006	0.018	Pass
12.361	12.358	-0.003	0.035	Pass
15.251	15.250	-0.001	0.025	Pass
21.908	21.911	0.003	0.036	Pass

**Table 7. Analysis of the z-axis angular rate**

Input %/s	Output %/s	Error %/s	Uncertainty %/s	Result
-24.512	-24.516	-0.005	0.040	Pass
-20.505	-20.507	-0.002	0.032	Pass
-16.147	-16.151	-0.004	0.029	Pass
-4.502	-4.506	-0.004	0.030	Pass
-0.022	-0.018	0.004	0.040	Pass
-0.006	-0.007	-0.001	0.027	Pass
-0.004	-0.011	-0.007	0.043	Pass
-0.001	-0.004	-0.003	0.021	Pass
0.001	0.009	0.008	0.036	Pass
0.003	0.002	-0.001	0.021	Pass
0.006	0.002	-0.004	0.035	Pass

Input %/s	Output %/s	Error %/s	Uncertainty %/s	Result
4.031	4.031	-0.000	0.042	Pass
16.912	16.919	0.007	0.027	Pass
21.723	21.728	0.005	0.031	Pass
27.976	27.975	-0.001	0.037	Pass

### 5.4. Summaries

Summaries of the preceding temperature, accelerometer and gyroscope results:

**Table 8. Summaries of calibration results**

Summary	Result
Temperature change result	Pass
Accelerometer x-axis result	Pass
Accelerometer y-axis result	Pass
Accelerometer z-axis result	Pass
Gyroscope x-axis result	Pass
Gyroscope y-axis result	Pass
Gyroscope z-axis result	Pass

### 6. Certificate Authorization

Authorization details regarding revision 1 of this calibration certificate:

Name [REDACTED]  
 Date 2023-01-11